DIN 58345: 2007

We, Vestfrost Solutions A/S declare under sole responsibility that the following products:

Product category: Model no.: Pharmacy Refrigerator 337, 397

Is incompliance with DIN 58345.

DIN stands for "Deutsches Institut für Normung" (German Institute for Standardization). The Institute's specialist bodies define industry standards. DIN 58345 relates to pharmaceutical refrigerators that are used to store pharmaceuticals at an operating temperature of 2°C to 8°C. DIN 58345 currently reflects the latest technological refrigerators.

Basic requirements established by DIN 58345:

- Operating temperature: +2°C to +8°C
- Usable in ambient temperature: +10°C to +35°C
- Noise emissions: < 60dB(A)
- Visual and acoustic warning system for high and low temperatures
- Dry contact for external alarm systems
- Warning system active from min. 12 h in the event of a power cut
- Safety thermostat for preventing temperatures below +2°C
- Load bearing capacity at least 100 kg/m2
- Lockable door

(1, Muls	10.04.2019
SIGNATURE	DATE



AUTHORITY: **TAJ BERNTH, COO** A/S VESTFROST 8181XXX rev 01